

**FRIEDMAN & BRUYA, INC.**

**ENVIRONMENTAL CHEMISTS**

Date of Report: 12/03/03

Date Received: 11/14/03

Project: Metro Self Monitor, PO# M92191, F&BI 311124

Date Extracted: 12/01/03

Date Analyzed: 12/01/03

**RESULTS FROM THE ANALYSIS OF THE WATER SAMPLE  
FOR TOTAL METALS BY  
INDUCTIVELY COUPLED PLASMA (ICP)  
(METHOD 6010)**

Results Reported as mg/L (ppm)

<u>Sample ID</u> Laboratory ID	<u>Chromium</u>	<u>Copper</u>	<u>Nickel</u>	<u>Zinc</u>
M92191 311124-01	0.77	0.69	0.84	<0.05
Method Blank	<0.05	<0.05	<0.10	<0.05

**FRIEDMAN & BRUYA, INC.****ENVIRONMENTAL CHEMISTS**

Date of Report: 12/03/03

Date Received: 11/14/03

Project: Metro Self Monitor, PO# M92191, F&amp;BI 311124

**QUALITY ASSURANCE RESULTS  
FROM TOTAL METALS BY  
INDUCTIVELY COUPLED PLASMA (ICP)  
(METHOD 6010)**

Laboratory Code: 311246-01 (Duplicate)

Analyte	Reporting Units	Sample Result	Duplicate Result	Relative Percent Difference	Acceptance Criteria
Chromium	mg/L (ppm)	0.21	0.21	0	0-20
Copper	mg/L (ppm)	0.16	0.16	0	0-20
Nickel	mg/L (ppm)	0.15	0.15	0	0-20
Zinc	mg/L (ppm)	0.05	<0.05	nm	0-20

Laboratory Code: 311246-01 (Matrix Spike)

Analyte	Reporting Units	Spike Level	Sample Result	% Recovery MS	Acceptance Criteria
Chromium	mg/L (ppm)	2	0.21	90	80-120
Copper	mg/L (ppm)	2	0.16	86	80-120
Nickel	mg/L (ppm)	4	0.15	85	80-120
Zinc	mg/L (ppm)	2	0.05	91	80-120

Laboratory Code: Laboratory Control Sample

Analyte	Reporting Units	Spike Level	% Recovery LCS	% Recovery LCSD	Acceptance Criteria	RPD (Limit 20)
Chromium	mg/L (ppm)	2	105	95	80-120	10
Copper	mg/L (ppm)	2	105	94	80-120	11
Nickel	mg/L (ppm)	4	106	97	80-120	9
Zinc	mg/L (ppm)	2	107	98	80-120	9

nm - The analyte was not detected in one or more of the duplicate analyses. Therefore, calculation of the RPD is not applicable.

311124

## SAMPLE CHAIN OF CUSTODY

CM 11/14/03

AT

Send Report To

G. Thompson

Company

ALASKAN Copper works

Address

628 S. Harbor St

City, State, ZIP

Seattle WA 98134

Phone #

206-382-8729

Fax #

SAMPLERS (signature)

PROJECT NAME/NO.

PO #

metno self monitor

m92191

REMARKS

Page # of

TURNAROUND TIME

☐ Standard (2 Weeks)☐ RUSH

Rush charges authorized by:

SAMPLE DISPOSAL

☐ Dispose after 30 days☐ Return samples☐ Will call with instructions

Sample ID	Lab ID	Date	Time	Sample Type	# of containers	ANALYSES REQUESTED										Notes	
						TPH-Diesel	TPH-Gasoline	BTEX by 8021B	VOCs by 8260	SVOCs by 8270	HFS						
M 92191	01	11/17/03	11:30	A20	1												

Friedman & Bruya, Inc.  
3012 16th Avenue West

Seattle, WA 98119-2029

Ph. (206) 285-8282

Fax (206) 283-5044

SIGNATURE

Relinquished by:

Received by:

Relinquished by:

Received by:

PRINT NAME

COMPANY

DATE

TIME

G. Thompson

Aew

11/14/03

11:30

Nhan Phan

Friedman &amp; Bruya

11/14/03

11:40

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.  
Charlene Morrow, M.S.  
Yelena Aravkina, M.S.  
Bradley T. Benson, B.S.  
Kurt Johnson, B.S.

3012 16th Avenue West  
Seattle, WA 98119-2029  
TEL: (206) 285-8282  
FAX: (206) 283-5044  
e-mail: fbi@isomedia.com

December 3, 2003

DUPLICATE COPY

INVOICE # 03ACU1203-4

Accounts Payable  
Alaskan Copper Works  
628 South Hanford  
Seattle, WA 98134

RE: Project Metro Self Monitor, PO# M92191, F&BI 311124 - Results of testing  
requested by Gerry Thompson for material submitted on November 14, 2003.

1 sample analyzed for Total Chromium, Copper, Nickel and Zinc by Method 6010B @ \$80 per sample	\$ 80.00
Amount Due .....	\$ 80.00

FEDERAL TAX ID # (b) (6)



FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.  
Charlene Morrow, M.S.  
Yelena Aravkina, M.S.  
Bradley T. Benson, B.S.  
Kurt Johnson, B.S.

3012 16th Avenue West  
Seattle, WA 98119-2029  
TEL: (206) 285-8282  
FAX: (206) 283-5044  
e-mail: fbi@isomedia.com

December 3, 2003

Gerry Thompson, Project Manager  
Alaskan Copper Works  
628 South Hanford  
Seattle, WA 98134

Dear Mr. Thompson:

Included are the results from the testing of material submitted on November 14, 2003 from the Metro Self Monitor, PO# M92191, F&BI 311124 project. There are 2 pages included in this report. Any samples that may remain are currently scheduled for disposal in 30 days. If you would like us to return your samples or arrange for long term storage at our offices, please contact us as soon as possible.

We appreciate this opportunity to be of service to you and hope you will call if you should have any questions.

Sincerely,

FRIEDMAN & BRUYA, INC.



Charlene Morrow  
Chemist

Enclosures  
ACU1203R.DOC